Technical Specifications EyeSeeCam vHIT





General Technical Specifications

Medical CE-mark:	The CE-mark indicates that Interacoustics A/S meets the requirements of Annex II of the Medical device Directive 93/42/EEC.	
	Approval of the quality system is made by TÜV – Identification no. 0123	
Standards:	Safety:	IEC 60601-1, Class II, Type B, applied parts
	EMC:	IEC 60601-1-2
Operating environment	Temperature:	15 - 35° C
	Relative Humidity:	30 – 90%
	Ambient Pressure:	98 kPa – 104 kPa
Transport & Storage	Storage Temperature:	10 - 35° C
	Transport Temperature:	10 - 50° C
	Relative Humidity:	10 – 95%

PC Interface	USB 2.0	
Camera	Monocular camera, interchangeable between left and right eye	
Goggle	Molded, flexible frame with snug fit, cleanable silicone contact areas. Fits pediatric and adult faces.	
Weight	Goggle: 40g	
	Camera: 32g	
	Combined weight: 72g	
Calibration	Goggle mounted laser with 5-point calibration	
Sample rate	220 Hz	
Head tracking sensor	Inertial Measurement Unit (IMU) gyroscope with 6 degrees of freedom	
Gain	Instantaneous gain at 40, 60 and 80 ms – equivalent to scleral search coil method.	
Data Analysis	Velocity Regression Plot, Gain Plot	

System Requirements:

PC Hardware	Intel Core i5 processor 2.53 GHz or better
	Minimum 4 GB RAM
	Hard drive with min. 100 GB space
	Min. display of 1280x720 pixels (1600 x 900 or higher recommended)
	DirectX 9.0C compatible graphics (Intel/NVidia recommended)
	One or more USB ports, version 2.0
PC Software	Windows®8 (32-bit or 64-bit)
	Windows®7 (32-bit or 64-bit)
	Windows® Media Player
	Adobe®Reader®





Included Parts

EyeSeeCam – vHIT camera and goggle	
USB cable	
EyeSeeCam vHIT Software and manuals CD	
OtoAccess Database software CD	
Ferrite Beads	

