





Hearing Instrument Restoration System

Designed to give dispensers & technicians a functional tool for repairing, restoring and maintaining hearing aids.

Effectively removes cerumen and debris from hearing aid tubing and ports.

UltraVac+

Technical Specs

Power Requirements: 100 - 240 V, 50/60 Hz

Power Consumption: <100 Watts

Fuse: 2.5A @ 100-240V

Operating Temperature: 0° To 50° C • 32° To 120° F

Debris Filter: I-63S

Vacuum Tips: #14, #15, 2-#18 and 2-#20 Gauge With Safety Lock Attachment

Drying Chamber Cycle Time: 4 Minutes With Auto Shut-off

Vacuum Wand: Continuous Use

Pressure Wand: Continuous Use

Dimensions: Approx. 29 cm x 31 cm x 16 cm (L x W x H) Approx. 11.25" x 12" x 6.25" (L x W x H)

Weight: Approx. 5 kg • Approx. 11 lbs

Standard Accessories: Drying Chamber Cup With Filter, Vacuum Wand With Filter, Pressure Wand, 6 Vacuum Wand Tips, Cleaning Tool Kit, Instruction Manual

MedRx ULTRA VAC

- Digital Display of Chamber Pressure and Time Remaining
- Automatic Detection of Operating Mode
- Separate Pressure and Vacuum Wands Minimize Debris in System
- In-Line Vacuum Filter for Added Reliability
- 4 Minute Drying Chamber Cycle
- Auto Shut Off

Functions of your UltraVac+

The Drying Chamber – Displaces moisture trapped within the hearing instruments. You can place two ITEs, or one BTE with an earmold, into the chamber at one time.

The Vacuum Wand – Used to extract ear wax and debris from the receiver tubes, microphone tubes, vents and battery compartments of the hearing instruments.

The Pressure Function – Used to clear obstructed vacuum wand tips and to blow debris from the receiver tubes, microphone tubes, vents and battery compartments of the hearing instruments.



MedRx International MAICO Diagnostics GmbH Sickingenstr. 70-71 10553 Berlin, Germany Phone: +49 30 70 71 46 50 Fax: +49 30 70 71 46 99 Email: medrx-sales@maico.biz www.medrx-diagnostics.com

Distributed by:	
MX-UV-INTDS-1 Rev. 1	



The Vacuum & Pressure Wands work simultaneously.