

Alta2 | Nera2 | Ria2

INSTRUCTIONS FOR USE

miniRITE



oticon
PEOPLE FIRST

Introduction to this booklet

This booklet guides you in how to use and maintain your new hearing instrument. Please read the booklet carefully including the **Warning section**. This will help you to achieve the full benefit of your new hearing instrument.

Your Hearing Care Professional has adjusted the hearing instrument to meet your needs. If you have additional questions, please contact your Hearing Care Professional.

| **About** | [Start up](#) | [Handling](#) | [Options](#) | [Warnings](#) | [Additional](#)s |

For your convenience this booklet contains a navigation bar to help you navigate easily through the different sections.

Indication for use

The hearing instrument is intended to amplify and transmit sound to the ear and thereby compensate for mild to moderate-to-severe hearing impairment.

IMPORTANT NOTICE

The hearing instrument amplification is uniquely adjusted and optimized to your personal hearing capabilities during the instrument fitting performed by your Hearing Care Professional.


Table of contents

About	Identify your hearing instrument speaker and ear piece	6
	miniRITE	8
	Identify left and right instrument	10
	Battery	11
Start up	Turn the hearing instrument ON and OFF	12
	When to replace a battery	13
	Replace the battery	14
Handling	Caring for your hearing instrument	16
	Put on the instrument	18
	Speaker with dome or GripTip	20
	Replace dome or GripTip	21
	Speaker with micro mold, LiteTip or power mold	22
	Maintain the micro mold, LiteTip or power mold	23


Options	Optional features and accessories	24
	Mute the hearing instrument (optional)	25
	Change programs (optional)	26
	Change volume (optional)	28
	Tamper-resistant battery drawer (optional)	30
	Wireless accessories (optional)	32
	Other options (optional)	33
Warnings	Warnings	34
Additional	Troubleshooting guide	44
	Water resistant	46
	Warranty certificate	48
	International warranty	49
	Mobile phone	50
	Technical information	52

Identify your hearing instrument speaker and ear piece

This will make it easier for you to navigate through this booklet.




- miniRITE with speaker**
A miniRITE comes with three different speakers and various ear pieces.



- miniRITE ear pieces**
A miniRITE will have one of the following ear pieces.




- miniRITE with Power mold speaker**




miniRITE

What it is




What it does




Identify left and right instrument

It is important to distinguish between the left and the right instrument as they might be programmed differently. You can find left/right color indicators on the battery drawer or on 60 and 85 speakers or on the molds.



Battery


Your hearing instrument is a miniature electronic device that runs on special batteries. To activate the hearing instrument, you must insert a new battery in the battery drawer. See how in the "Replace the battery" section.




Turn the hearing instrument ON and OFF

The battery drawer is also used to switch the hearing instrument on and off. To preserve the battery, make sure your instrument is switched off when you are not wearing it.

Turn ON
Close the battery drawer with the battery in place.



Turn OFF
Open the battery drawer



When to replace a battery

When it is time to replace the battery you will hear two beeps repeated in moderate intervals until the battery runs out.



Two beeps
= The battery is running low




Four beeps
= The battery has run out

Battery maintenance tip

To make sure the hearing instrument is always working, bring spare batteries with you, or replace the battery before you leave home.

Replace the battery

1. Remove




Fully open the battery drawer. Remove the battery.

2. Uncover




Remove the sticky label from the + side of the new battery.

3. Insert




Insert the new battery into the battery drawer. Make sure the + side faces up.

4. Close



Close the battery drawer. The instrument will play a jingle through the ear piece. Hold the earpiece close to your ear to hear the jingle.

Tip



The MultiTool can be used for battery change. Use the magnetic end to remove and insert batteries.


The MultiTool is provided by your Hearing Care Professional.

Caring for your hearing instrument

When handling your hearing instrument, hold it over a soft surface to avoid damage if you drop it.

Clean the microphone openings

Carefully brush away debris from the openings. Gently brush the surface. Make sure the brush is clean and that it is not pressed into the openings.




IMPORTANT NOTICE

Use a soft, dry cloth to clean the hearing instrument. It must never be washed or immersed in water or other liquids.

The MultiTool


The MultiTool contains a brush and a wire loop for cleaning ear wax from the mold. The brush can be replaced and purchased from your Hearing Care Professional.



Put on the instrument

The speaker brings the sound into your ear. The speaker should always be used with an ear piece attached. Use only parts designed for your hearing instrument.

If the speaker has an ear grip, place it in the ear so it follows the contour of the ear (see step 3).




Step 1



Place the hearing instrument behind your ear.

Step 2



Hold the bend of the speaker wire between your thumb and index finger. The ear piece should point towards the ear canal opening.


Step 3



Gently push the ear piece into your ear canal until the speaker wire sits close against the side of your head.

Speaker with dome or GripTip

The dome and Grip Tip are made from soft, medically approved rubber material. There are 4 different types of domes. Check your dome type and size below.




IMPORTANT NOTICE

If the dome is not on the tube when removed from the ear, the dome might still be in the ear canal. Consult your Hearing Care Professional for further instructions.

Replace dome or GripTip

The dome or GripTip should not be cleaned. When it is more or less filled with wax, replace it with a new one.



Speaker with micro mold, LiteTip or power mold

There are 3 different types of molds: micro mold, LiteTip and power mold. The molds are customized for your ear shape.



• Micro mold



• LiteTip




• Power mold

Maintain the micro mold, LiteTip or power mold

The mold should be cleaned regularly.

The wax filter in the ear piece should be replaced when clogged or when the hearing instrument does not sound normal.

Use the wax filter delivered by your Hearing Care Professional.



Optional features and accessories


The features and accessories described on the following pages are optional. Please contact your Hearing Care Professional to find out how your hearing instrument is programmed.

If you experience difficult listening situations, a special program may be helpful. These are programmed by your Hearing Care Professional.

Write down hearing situations where you may need help.

Mute the hearing instrument (optional)

Use the mute function if you need to silence the instrument while wearing it.




- Apply a very long press to the button to mute the instrument. To reactivate the instrument, push the button briefly.

IMPORTANT NOTICE

Do not use the mute function as an off switch, as the hearing instrument still draws current from the battery in this mode.

Change programs (optional)





Your hearing instrument can have up to 4 different programs. These are programmed by your Hearing Care Professional.



• Press the button to change program

Note that if you have two instruments, the RIGHT instrument switches forward from e.g. program 1 to 2 and the LEFT instrument switches backwards from e.g. program 4 to 3.

To be filled out by the Hearing Care Professional

Program	Sound you will hear when activated	When to use
1	 "1 beep"	
2	 "2 beeps"	
3	 "3 beeps"	
4	 "4 beeps"	

Program change:

LEFT


RIGHT

Short press

Long press

□ **Change volume** (optional)


When you have two instruments, the push button allows you to adjust the volume in both ears. You will hear a click when you turn the volume up or down.




Quick reset

If you wish to return to the standard settings of the instrument programmed by your Hearing Care Professional, simply open and then close the battery drawer.

Open



Close



□ **Tamper-resistant battery drawer** (optional)


To keep the battery out of reach of infants, small children and people with learning difficulties, a tamper-resistant battery drawer should be used. Use a small screwdriver to open the drawer.

IMPORTANT NOTICE

Avoid using excessive force to open the battery drawer. Do not force the battery drawer beyond its fully opened position. Make sure to insert the battery correctly.

Please contact your Hearing Care Professional if you suspect the tamper-resistant effectiveness and locking ability is damaged.

Open and close the battery drawer



Insert the screwdriver into the slit on the back of the hearing instrument. Close the drawer using your fingers. Make sure the drawer is completely closed.

Wireless accessories (optional)

As an enhancement to your wireless hearing instrument, a broad range of wireless accessories are available.

ConnectLine

ConnectLine is a family of products that allows you to receive audio signals from TVs, phones, music players, PCs or an external microphone wirelessly through your hearing instrument.

Remote Control

The Remote Control offers an opportunity to change program or to adjust the volume in your hearing instrument.

Other options (optional)



Autophone

is a program that can be automatically activated if your telephone has a built-in teleloop. A magnet may need to be placed on your telephone next to the sound outlet.

Warnings

You should familiarize yourself fully with the following general warnings before using your hearing instrument for personal safety and to secure correct use.

Consult your Hearing Care Professional if you experience unexpected operations or events with your hearing instrument.

Usage of hearing instruments

- Hearing instruments should be used only as directed and adjusted by your Hearing Care Professional. Misuse can result in sudden and permanent hearing loss.

- Never allow others to wear your hearing instrument as incorrect usage could cause permanent damage to their hearing.

Choking hazards & risk of swallowing batteries

- Hearing instruments, their parts, and batteries should be kept out of reach of children and anyone who might swallow these items, or otherwise cause injury to themselves.

- Batteries have occasionally been mistaken for pills. Therefore check your medicine carefully before swallowing any pills.
- Most hearing instruments can be supplied with a tamper-resistant battery drawer upon request. This is strongly recommended for infants, small children, and people with learning difficulties.
- Children below 36 months must always use a tamper-resistant battery drawer. This option is available in BTE, miniBTE, RITE and miniRITE instruments.

If a battery or hearing instrument is swallowed, see a doctor immediately and contact the National Poison Center at 1-800-222-1222 or National Battery Ingestion Hotline at 202-625-3333.

Battery use

- Always use batteries recommended by your Hearing Care Professional. Batteries of low quality may leak and cause bodily harm.
- Never attempt to recharge your batteries and never dispose of batteries by burning them. There is a risk that the batteries will explode.

Warnings

Dysfunction

- Be aware of the possibility that your hearing instrument may stop working without notice. Keep this in mind when you depend on warning sounds (e.g., when you are in traffic). The hearing instruments may stop functioning, for instance if the batteries have expired or if the tubing is blocked by moisture or ear wax.

Active implants

- Caution must be taken with active implants. In general, follow the guidelines recommended by manufacturers of implantable defibrillators and pacemakers regarding use with mobile phones.
- If you wear an active implant, then keep the hearing instrument more than 15 cm away from the implant. If you have an Autophone magnet or MultiTool (which has a built-in magnet), keep them more than 30 cm away from the implant. E.g., do not carry them in a breast pocket.

- If you have an active brain implant, please contact the manufacturer of your implantable device for information about the risk of disturbance.

Explosives

- The power source in your hearing instrument has insufficient energy to cause fire in normal usage conditions. The hearing instrument has not been tested for compliance with international standards concerning explosive environments. It is recommended not to use your hearing instrument in areas where there is a danger of explosions.

X-ray, CT, MR, PET scanning and electrotherapy

- Remove your hearing instrument for example during X-ray, CT / MR / PET scanning electrotherapy or surgery as your hearing instrument may be damaged when exposed to strong fields.

Warnings

Avoiding heat and chemicals

- Your hearing instrument must never be exposed to extreme heat e.g., left inside a parked car in the sun.
- Your hearing instrument must not be dried in microwave ovens or other ovens.
- The chemicals in cosmetics, hairspray, perfume, after shave lotion, suntan lotion and insect repellent can damage your hearing instrument. Always remove your hearing instrument before applying such products and allow time to dry before putting it on.

Power instrument

- Special care should be exercised in selecting, fitting and using a hearing instrument where maximum sound pressure capability exceeds 132 dB SPL (IEC 711), as there may be risk of impairing the remaining hearing of the hearing instrument user.

For information on whether your instrument is a power instrument, speak with your hearing care professional or consult the settings overview page at the end of this manual.

Possible side effects

- Hearing instruments, molds or domes may cause an accelerated accumulation of ear wax.
- The otherwise non-allergenic materials used in hearing instruments may in rare cases cause a skin irritation or any other unusual condition.

Please seek consultation with a physician if these conditions occur.

Interference

- Your hearing instrument has been thoroughly tested for interference, according to the most stringent international standards. However, interference with your hearing instrument and other devices may occur, (e.g., some mobile telephones, citizens band systems and shop alarm systems). If this occurs, increase the distance between the hearing instrument and the device.

Warnings

Warning to hearing instrument dispensers

A hearing instrument dispenser should advise a prospective hearing instrument user to consult immediately with a licensed physician (preferably an ear specialist) before dispensing a hearing instrument if the hearing instrument dispenser determines through inquiry, actual observation, or review of any other available information concerning the prospective user, that the prospective user has any of the following conditions:

- (i) Visible congenital or traumatic deformity of the ear.
- (ii) History of active drainage from the ear within the previous 90 days.

- (iii) History of sudden or rapidly progressive hearing loss within the previous 90 days.
- (iv) Acute or chronic dizziness.
- (v) Unilateral hearing loss of sudden or recent onset within the previous 90 days.
- (vi) Audiometric air-bone gap equal to or greater than 15 decibels at 500 Hertz (Hz), 1,000 Hz, and 2,000 Hz.
- (vii) Visible evidence of significant cerumen accumulation or a foreign body in the ear canal.
- (viii) Pain or discomfort in the ear.

Special care should be exercised in selecting and fitting a hearing instrument whose maximum sound pressure capability exceeds 132 dB SPL as there may be risk of impairing the remaining hearing of the hearing instrument user.

Important notice for prospective hearing instrument users

- Good health practice requires that a person with a hearing loss have a medical evaluation by a licensed physician (preferably a physician who specializes in diseases of the ear) before purchasing a hearing instrument. Licensed physicians who specialize in diseases of the ear are often referred to as

Otolaryngologists, Otologists or Otorhinolaryngologists. The purpose of medical evaluation is to ensure that all medically treatable conditions that may affect hearing are identified and treated before the hearing instrument is purchased. Following the medical evaluation, the physician will give you a written statement that states that your hearing loss has been medically evaluated and that you may be considered a candidate for a hearing instrument. The physician will refer you to an audiologist or a hearing instrument dispenser, as appropriate, for a hearing instrument evaluation.

Warnings

- The audiologist or hearing instrument dispenser will conduct a hearing instrument evaluation to assess your ability to hear with and without a hearing instrument. The hearing instrument evaluation will enable the audiologist or dispenser to select and fit a hearing instrument to your individual needs. If you have reservations about your ability to adapt to amplification, you should inquire about the availability of a trial, rental or purchase-option program. Many hearing instrument dispensers now offer programs that permit you to wear a hearing instrument for a period of time for a nominal fee, after which you may decide if you want to purchase the hearing instrument. Federal law limits the sale of hearing instruments to those individuals who have obtained a medical evaluation from a licensed physician.
 - Federal law permits a fully informed adult to sign a waiver statement declining the medical evaluation for religious or personal beliefs that preclude consultation with a physician. The exercise of such a waiver is not in your best health interest and its use is strongly discouraged. A hearing instrument will not restore normal hearing and will not prevent or improve a hearing impairment resulting from organic conditions. A hearing instrument is only part of hearing rehabilitation and may need to be supplemented by auditory training and lip reading.
- Children with hearing loss**
In addition to seeing a physician for medical evaluation, a child with a hearing loss should be directed to an audiologist for evaluation and rehabilitation, since hearing loss may cause problems in language development and educational and social growth of a child. An audiologist is qualified by training and experience to assist in the evaluation and rehabilitation of a child with a hearing loss. If the user is an infant, small child, or person of mental incapacity, it is recommended that the hearing instrument be modified with a tamper-resistant battery compartment.

Troubleshooting guide

Symptom	Possible causes	Solutions
No sound	Worn-out battery	Replace the battery
	Clogged tube or earpieces (mold, dome, GripTip, micro mold or LiteTip)	Clean mold
		Consider replacing wax filter or dome
Intermittent or reduced sound	Clogged sound outlet	Clean mold or replace wax filter or dome
	Moisture	Wipe battery with a dry cloth
	Worn-out battery	Replace the battery
Squealing noise	Hearing instrument earpiece inserted incorrectly	Re-insert hearing instrument
	Ear wax accumulated in ear canal	Have ear canal examined by your doctor

If none of the above solutions work, consult your Hearing Care Professional for assistance.

Water resistant

Your hearing instrument is water resistant which means it is designed to be worn in all daily life situations. Therefore you do not have to worry about sweat or getting wet in the rain. Should your hearing instrument come in contact with water and stop working, please follow these guidelines:

1. Gently wipe off any water
2. Open the battery drawer and remove the battery and gently wipe off any water in the battery drawer
3. Let the hearing instrument dry with the battery drawer left open for approximately 30 minutes
4. Insert a new battery

IMPORTANT NOTICE

Do not wear your hearing instrument while showering or participating in water activities. Do not immerse your hearing instrument in water or other liquids.

Warranty certificate

Name of owner: _____

Dispenser: _____

Dispenser address: _____

Dispenser phone: _____

Purchase date: _____

Warranty period: _____ Month: _____

Model left: _____ Serial no.: _____

Model right: _____ Serial no.: _____

Battery size: _____

International warranty

Your hearing instrument is covered by an international limited warranty issued by the manufacturer for a period of 12 months from the date of delivery. This limited warranty covers manufacturing and material defects in the hearing instrument itself, but not accessories such as batteries, tubing, ear wax filters, etc. Problems arising from improper handling or care, excessive use, accidents, repairs made by an unauthorized party, exposure to corrosive conditions, physical changes in your ear, damage due to foreign objects entering the device, or incorrect adjustments are NOT covered by the limited warranty and may void it. The above warranty does not affect any legal rights that you

might have under applicable national legislation governing sale of consumer goods. Your Hearing Care Professional may have issued a warranty that goes beyond the clauses of this limited warranty. Please consult him/her for further information.

If you need service

Take your hearing instrument to your Hearing Care Professional, who may be able to sort out minor problems and adjustments immediately.

Mobile phone

Some hearing instrument users have reported a buzzing sound in their hearing instrument when they are using mobile phones, indicating that the mobile phone and hearing instrument may not be compatible. According to the ANSI C63.19 standard (ANSI C63.19-2007 American National Standard Methods of Measurement of Compatibility Between Wireless Communications Devices and Hearing Aids), the compatibility of a particular hearing instrument and mobile phone can be predicted by adding the rating for the hearing instrument immunity to the rating for the mobile phone emissions. For example, the sum of a

hearing instrument rating of 2 (M2/T2) and a telephone rating of 3 (M3/T3) would result in a combined rating of 5. Any combined rating that equals at least 5 would provide “normal use”; a combined rating of 6 or greater would indicate “excellent performance”.

The immunity of this hearing instrument is M4. The equipment performance measurements, categories, and system classifications are based upon the best information available, but cannot guarantee that all users will be satisfied.

IMPORTANT NOTICE

The performance of individual hearing instruments may vary with individual mobile phones. Therefore, please try this hearing instrument with your mobile phone or, if you are purchasing a new phone, be sure to try it with your hearing instrument prior to purchase.

For additional guidance, please ask your mobile phone provider for the booklet entitled “Hearing Aid Compatibility with Digital Wireless Mobile Phones.”

Technical information

The hearing instrument contains a radio transmitter using short range magnetic induction technology working at 3.84 MHz. The magnetic field strength of the transmitter is < -42 dB μ A/m @ 10m.

The emission power from the radio system is well below international emission limits for human exposure. For comparison, the radiation of the hearing instrument is lower than unintended electromagnetic radiation from, for example, halogen lamps, computer monitors, dishwashers, etc. The hearing instrument complies with international standards concerning Electromagnetic Compatibility.

Due to the limited space available on the instruments, all relevant approval markings are found in this document.

The hearing instrument contains a module with:

FCC ID: U28FU2MRIT
IC: 1350B-FU2MRIT

The device complies with Part 15 of the FCC rules and RSS-210 of Industry Canada.

Operation is subject to the following two conditions:

1. This device may not cause harmful interference.
2. This device must accept any interference received, including interference that may cause undesired operation.

Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

More information: www.oticon.com

Oticon declares that this hearing instrument is in compliance with the essential requirements and other relevant provisions of Directive 1999/5/EC.

Declaration of conformity is available at:

Oticon A/S
Kongebakken 9
DK-2765 Smørum
Denmark
www.oticon.com

CE 0543 0682



Waste from electronic equipment must be handled according to local regulations.



N1175

IP58

Power instrument <input type="checkbox"/> Yes <input type="checkbox"/> No				
Settings overview for your hearing instrument				
Left			Right	
<input type="checkbox"/> Yes	<input type="checkbox"/> No	Volume control	<input type="checkbox"/> Yes	<input type="checkbox"/> No
<input type="checkbox"/> Yes	<input type="checkbox"/> No	Program shift	<input type="checkbox"/> Yes	<input type="checkbox"/> No
<input type="checkbox"/> Yes	<input type="checkbox"/> No	Mute	<input type="checkbox"/> Yes	<input type="checkbox"/> No
Volume control indicators				
<input type="checkbox"/> On	<input type="checkbox"/> Off	Beeps at min /max volume	<input type="checkbox"/> On	<input type="checkbox"/> Off
<input type="checkbox"/> On	<input type="checkbox"/> Off	Clicks when changing volume	<input type="checkbox"/> On	<input type="checkbox"/> Off
<input type="checkbox"/> On	<input type="checkbox"/> Off	Beeps at preferred volume	<input type="checkbox"/> On	<input type="checkbox"/> Off
Battery indicators				
<input type="checkbox"/> On	<input type="checkbox"/> Off	Low battery warning	<input type="checkbox"/> On	<input type="checkbox"/> Off

TECHNICAL DATA

miniRITE

Alta2

oticon
PEOPLE FIRST

Measurements according to American National Standard ANSI S3.22 (2003) and S3.7 (1995)

Supply voltage:
Battery Zinc
Air 1.4 Volt

0 dB SPL ref. 20 μ Pa

miniRITE 60

60

Peak OSPL90	105 dB SPL
HF Average OSPL90	103 dB SPL
Peak Full-on Gain	35 dB
HF Average Full-on Gain	30 dB
Reference Test Gain	26 dB
Frequency Range	100-8300 Hz
Total Harmonic Distortion 500 Hz	<2%
Total Harmonic Distortion 800 Hz	<2%
Total Harmonic Distortion 1600 Hz	<2%
Battery Consumption	1.3 mA
Equivalent Input Noise Level (omni/dir)	16/24 dB SPL
HF Average SPLITS (left/right ear)	-
Attack Time	1 ms
Release Time	120 ms

miniRite 85

85

118 dB SPL	124 dB SPL	125 dB SPL
114 dB SPL	121 dB SPL	121 dB SPL
55 dB	57 dB	61 dB
47 dB	52 dB	55 dB
38 dB	44 dB	44 dB
100-8700 Hz	100-8100 Hz	100-7900 Hz
<2%	<2%	<2%
<2%	<2%	<2%
<2%	<2%	<2%
1.2 mA	1.3 mA	1.3 mA
18/25 dB SPL	16/25 dB SPL	16/28 dB SPL
-	-	-
1 ms	1 ms	1 ms
60 ms	30 ms	123 ms

miniRite 100

100

miniRITE 105


105

OSPL90 - Output Sound Pressure Level

Input: 90 dB SPL.
Technical setting: A0


miniRITE 60

60




miniRITE 85

85




miniRITE 100

100





miniRITE 105

105



Full-on Gain

Input: 50 dB SPL.
Technical setting: A0





OSPL90 - Output Sound Pressure Level

Input: 90 dB SPL.
Technical setting: A0

Full-on Gain

Input: 50 dB SPL.
Technical setting: A0



TECHNICAL DATA

miniRITE

Nera2, Ria2

oticon
PEOPLE FIRST

Measurements according to American National Standard ANSI S3.22 (2003) and S3.7 (1995)

Supply voltage:
Battery Zinc
Air 1.4 Volt

0 dB SPL ref. 20 μ Pa

miniRite 60

60

Peak OSPL90	105 dB SPL
HF Average OSPL90	103 dB SPL
Peak Full-on Gain	35 dB
HF Average Full-on Gain	30 dB
Reference Test Gain	26 dB
Frequency Range	100-7000 Hz
Total Harmonic Distortion 500 Hz	<2%
Total Harmonic Distortion 800 Hz	<2%
Total Harmonic Distortion 1600 Hz	<2%
Battery Consumption	1.3 mA
Equivalent Input Noise Level (omni/dir)	16/24 dB SPL
HF Average SPLITS (left/right ear)	-
Attack Time	1 ms
Release Time	120 ms

miniRite 85

85

118 dB SPL	124 dB SPL	125 dB SPL
114 dB SPL	121 dB SPL	121 dB SPL
55 dB	57 dB	61 dB
47 dB	52 dB	55 dB
38 dB	44 dB	44 dB
100-7200 Hz	100-7200 Hz	100-6900 Hz
<2%	<2%	<2%
<2%	<2%	<2%
<2%	<2%	<2%
1.2 mA	1.3 mA	1.3 mA
18/25 dB SPL	16/25 dB SPL	16/28 dB SPL
-	-	-
1 ms	1 ms	1 ms
60 ms	30 ms	123 ms

miniRite 100

100


miniRITE 105

OSPL90 - Output Sound Pressure Level

Input: 90 dB SPL.
Technical setting: A0


miniRITE 60

60



miniRITE 85

85




151781US / 01.15

OSPL90 - Output Sound Pressure Level

Input: 90 dB SPL.
Technical setting: A0


miniRITE 100

100



miniRITE 105



105



151781US / 01.15



Full-on Gain

Input: 50 dB SPL.
Technical setting: A0



Full-on Gain

Input: 50 dB SPL.
Technical setting: A0



People First

People First is our promise
to empower people
to communicate freely,
interact naturally and
participate actively



0000151781000001

oticon
PEOPLE FIRST