
Technical data sheet

Oticon Siya 1 & 2



90

		Oticon Siya 1	Oticon Siya 2	
Speech Understanding	Noise Reduction LX	•	•	
	Multiband Adaptive Directionality LX	•	•	
	Single Compression LX	•	•	
	Speech Rescue™ LX	•	-	
Sound Quality	Fitting Bandwidth*	8 KHz	8 KHz	
	Processing Channels	48	48	
	Bass Boost (streaming)	•	•	
Listening Comfort	Transient Noise Management	On/Off	-	
	Feedback shield LX	•	•	
	Wind Noise Management	•	•	
	Binaural Coordination***	•	•	
Optimising Fitting	Fitting Bands	10	8	
	Adaptation Management	•	•	
	Oticon Firmware Updater	•	•	
	Multiple Directionality options	•	•	
	Fitting Formulas	NAL-NL1+2, DSL v5.0	NAL-NL1+2, DSL v5.0	
Connecting to the World	Stereo streaming (2.4 GHz)	0	0	
	Oticon ON App	0	0	
	ConnectClip	0	0	
	Remote Control 3.0	0	0	
	TV Adapter 3.0	0	0	
	Tinnitus SoundSupport™***	0	0	
	Battery life, hours**	55-60/105-115	55-60/105-115	

Bandwidth accessible for gain adjustments during fitting

Battery size 312 - IEC PR41 / Battery size 13 - IEC PR48.

Real usage battery life is shown as an estimated interval based on mixed use cases with variable amplification settings and variable input levels, incl. direct stereo streaming from a TV (25% of the time) and streaming from a mobile phone (6% of the time)

*** If push button is choosen

• Default

o Optional

Not included

OTICON | Siya ITC, ITE HS & FS 90



Oticon Siya ITC, ITE HS & FS introduce an updated faceplate design.

Oticon Siya is built on the powerful Velox™ platform, processing sound in 48 channels for highresolution sound quality.

Oticon Siya is a Made for iPhone® hearing aid that offers a full connectivity package built with 2.4 GHz Bluetooth for advanced and streamer free connectivity.

Fully programmable with updatable firmware, the Velox platform is ready for the future.




Technical data Measured according to	Ear Simulator IEC 60118-0:1983/AMD1:1994, IEC 60118-0:2015, IEC 60118-1:1995+AMD1:1998 CSV and IEC 60318-4:2010		2CC Coupler ANSI S3.22-2014, IEC 60118-0:2015 and IEC 60318-5:2006		
Oticon Siya ITC ITE HS & FS 90	Siya 1	Siya 2	Siya 1	Siya 2	
Frequency range Hz		110-7500		100-7500	
	Peak	131 dB SPL		120 dB SPL	
OSPL90	1600 Hz	123 dB SPL		115 dB SPL	
	HFA-OSPL90	124 dB SPL		116 dB SPL	
	Peak	65 dB		55 dB	
Full-on gain*	1600 Hz	57 dB		48 dB	
	HFA-FOG	58 dB		50 dB	
Reference test gain		48 dB		39 dB	
	1 mA/m field	87 dB SPL		-	
Telecoil output (1600 Hz)	10 mA/m field	107 dB SPL		-	
	SPLITS L/R	-		96/96 dB SPL	
T + 11	500 Hz	2%		<2%	
Total harmonic distortion (Input 70 dB SPL)	800 Hz	2 %		<2%	
(mput 70 db 31 L)	1600 Hz	2.0	%	<2	. %
Equivalent input poice level	Omni	18 dB SPL		15 dB SPL	
Equivalent input noise level	Dir	28 dB SPL		27 dB SPL	
Pattory consumption**	Typical	1.8 mA		1.8 mA	
Battery consumption**	Quiescent	1.7 mA		1.7 mA	
Battery life, calculated, hours 312 and 13***	100/175 100/17		/170		
IRIL (IEC 60118-13:2016)		700/1400/2000 MHz: 20/12/6 dB SPL			

Measured with the gain control of the hearing aid set to its full-on position minus 20 dB and with an input SPL of 70 dB. This is to obtain a gain response equal to the full-on gain response from e.g. IEC 60118-0+A1:1994 but without influence of feedback.

Battery current is measured according to IEC 60118-0:1983/AMD1:1994 §7.11, IEC 60118-0:2015 §7.7 and ANSI S3.22:2014 §6.13 after a settling time of minimum 3 minutes.

Based on the standardised battery consumption measurement (IEC 60118-0:1983/AMD1:1994). The actual battery life depends on battery quality, use pattern, active feature set, hearing loss and sound environment.



Technical information: Omnidirectional mode is used unless otherwise stated.

Operating conditions Temperature: +1°C to +40°C Relative humidity: 5% to 93%, non-condensing	Storage and transportation conditions Temperature and humidity should not exceed the following limits for extended periods during transportation and storage. Temperature: -25°C to +60°C Relative humidity: 5% to 93%, non-condensing
----------------------------------------------------------------------------------------------	-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------